



WEB NOTIFICATION SYSTEM

# Site Check

Gayles Telephone Box, Slip Inn Bank, North Yorkshire, DL11 7JE ((21)X19A099875)

**Device type:** AED+ (Zoll)

**Ambulance group:** Yorkshire Ambulance Service

**Serial number:** (21)X19A099875

**Ambulance site ID :** NCPAD1172

**Date:** Wednesday, July 17, 2024 - 11:21

**Reporter:** Ian Goulbourne

**Source:** 85.255.233.154

## SUBMITTED VALUES

Does the defibrillator have a cabinet?	Yes
Is the cabinet door operational?	Visually inspected and not opened (Rotaid Cabinet only)
Do all the lock buttons press okay?	Not applicable (unlocked cabinet)
Is there any sign of water in the cabinet or bag?	OK - clear
Is the light working?	Yes - all OK
Is the thermostat set for 7+ degrees?	Everything OK
Ready indicator status	OK
Is your battery inserted and has power?	Yes - All OK
Are the main pads connected and in date?	Yes
Is the rescue / ready kit with the defib?	Yes - All OK
Frequency of checks	Monthly

Cabinet status

ok

---

© COPYRIGHT 2009-2024 THE COMMUNITY HEARTBEAT TRUST  
TELEPHONE : 0330 1243067